



**ResMed**

# Declaration of Conformity



**Manufacturer:**

**ResMed Ltd**

1 Elizabeth Macarthur Drive  
Bella Vista  
NSW 2153  
Australia

**European Representative:**

**ResMed (UK) Ltd**

96 Milton Park  
Abingdon  
Oxfordshire OX14 4RY  
United Kingdom

**Product / Device Information**

Product Type:	GPRS Communication System
Trade Name(s):	AirCurve 10 Series
Model Number(s):	<b>370XX</b>

**Standard(s) applied**

Health (R&TTE, Article 3.1a)	EN 62311:2008	Full Compliance
Safety (R&TTE, Article 3.1a)	EN 60950-1:2006+ A11:2009+A12:2011	Full Compliance
EMC (R&TTE, Article 3.1b)	EN 301 489-1 V1.9.2 EN 301 489-7 V1.3.1	Full Compliance
Radio spectrum (R&TTE, Article 3.2)	EN 301 511 V9.0.2	Full Compliance

**Additional Information**

Notified Body Involved (Testing Organisation):	GPRS Communication System
Technical Construction file: Held by :	<b>ResMed Ltd</b> 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 <b>Australia</b>
R&TTE Radio Class:	GSM Class 4 (2W, 33 dBm) @ GSM 850/900 GSM Class 1 (1W, 30 dBm) @ GSM 1800/1900

We herewith declare that the products meet national law of the essential requirements and other relevant requirements of the Radio and Telecommunications Terminal Equipment (R&TTE Directive 1999/5/EC of 9. March 1999).

Signed at Sydney, Australia on: .....1<sup>st</sup> Oct 2014.....

GREG VOELKER FOR SIMON LEWIS

Dr. Simon Lewi  
Director – Regulatory Affairs  
ResMed Ltd