RESMED

Swift™ LT

NASAL PILLOWS SYSTEM

Fitting Guide

- Holding the pillows at your nose, pull the headgear over your head.
- Headgear should sit midway between the ears and eyes.

- Adjust the top straps by pulling them evenly until comfortable.
- Do not overtighten.

- Repeat Step 2 with the back straps.
- Do not overtighten.

- Using the click-fit adjustor, position the nasal pillows at the most suitable angle.
- Connect the Swift LT short tube to the device air tubing.
- Turn on the device.

- The Swift LT is now fitted and ready for use.

Optional

- For alternate tubing positions, use the tube retainer.

- To attach, push it into the top headgear buckle.
- Clip in either the short tube or device air tubing, depending on flexibility needs.
- The tubing can be positioned to either side of your face or centered.
- If the pillows pull on your nose, adjust the tubing at the tube retainer.

⚠ For details, refer to the *Swift LT Nasal Pillows System User Guide.*

ResMed Corp Poway, CA, USA +1 858 746 2400 or 1 800 424 0737 (toll free), ResMed Pty Ltd Bella Vista, NSW, Australia +61 (2) 8884 1000 or 1 800 658 189 (toll free). Offices in Austria, Brazil, China, Finland, France, Germany, Hong Kong, India, Japan, Malaysia, Netherlands, New Zealand, Norway, Singapore, Spain, Sweden, Switzerland, United Kingdom (see website for details). Protected by patents: AU 785376, DE 20321468.4, DE 20321469.2, DE 20321470.6, DE 20321471.4, DE 20321472.2, EP 1314445, NZ 553756, NZ 553757, NZ 553822, NZ 553824, NZ 553825, US 6581594, US 6823865, US 7159587, US 7318437. Other patents pending. Protected by design registrations: AU 301943, CN 200430102469.3, JP 1243026, NZ 407161, US D552731. Others designs pending. Swift is a trademark of ResMed Pty Ltd and Swift is registered in U.S. Patent and Trademark Office.
© 2008 ResMed. Specifications may change without notice. 1012483/1 08 03