



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Product:

Standard Trolley and Homecare Stand

Intended Use:

The ResMed Standard Trolley and Homecare Stand are designed for practical and convenient use to hold and move ResMed therapy devices and accessories between points of therapy in the home or hospital environment.

They are intended to support the ventilation device, humidifier, battery pack, oxygen cylinder, circuit support arm, and other clinical devices as needed.

Classification:

I according to Rule 1

GMDN:

42514 Ventilator stand

Conformity Assessment Route: Annex VII, 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

Signed at Sydney, Australia on: 11 December 2019

A handwritten signature in black ink, appearing to read "Johanna Wright".

Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC112

First issued: 19 February 2010