



Rx Only
English

Non-Vented Mask/Device compatibility list

This document specifies the masks compatible with ResMed's ventilators. For full details regarding the correct use of these devices including settings, please refer to the specific user guide.

Non-vented masks can be used with an active-exhaust-valve ventilator. The ventilators in this list all include an active exhaust valve.

Note: Not all products are available in all regions.

Device	AcuCare F1-0 / F1-1	Quattro Air NV
Astral™		
Astral 100	✓	✓
Astral 150	✓	✓
VS Ultra™		
VS Ultra	✓	✓
VS Integra	✓	✓
VS III™		
VS III	N/V	✓

N/V = The mask has not been validated for this device.



VS III **ResMed Paris** 240 rue de la Motte 77550 Moissy-Cramayel France
Astral **ResMed Ltd** 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia

DISTRIBUTED BY

ResMed Ltd 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia
ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA
ResMed (UK) (for Ltd) 96 Jubilee Ave Milton Park Abingdon Oxfordshire OX14 4RW UK

See www.resmed.com for other ResMed locations worldwide. For patent information, see www.resmed.com/ip. AcuCare, Quattro Air is a trademark of ResMed Ltd. Astral and VS III are trademarks of ResMed Paris SAS. Quattro Air and Astral are registered in U.S. Patent and Trademark Office. © 2015 ResMed Ltd. 1017302/4 2015-02